

Instrukcja używania
Instructions for use

COLLAR STARK HALO

ORTEZA CZASZKOWA / HALO BRACE

Numer katalogowy / Catalogue Number:	MQA4L
Basic UDI-DI:	59017804MQA4LME



MEYRA
medical



Producent/Manufacturer:

Mdh sp. z o.o.,
Ul. W. Tymienieckiego 22/24,
90-349 Łódź
Tel. +48 426 748 384
www.mdh.pl, www.qmedinfo.pl

Wydanie 11.05.2021
Edition 11.05.2021

INSTRUKCJA UŻYWANIA

COLLAR STARK HALO – Orteza czaszkowa

Orteza Czaszkowa COLLAR STARK HALO dla dorosłych jest przystosowany do zastosowania wyłącznie z kołnierzem COLLAR STARK. Właściwe stosowanie stabilizatora ma istotne znaczenie dla unieruchomienia i komfortu pacjenta. Dwuczęściowa konstrukcja składa się z części przedniej i części tylnej, które są pakowane w zestawie.

ZAKŁADANIE:

1. Dopasuj Stabilizator Halo do plastikowego wzmocnienia w tylnej części kołnierza przez wsunięcie od góry i zaczepienie stabilizatora o element podpierający potylicę. Przesuń obie części tak aby były równe.
2. Po prawidłowym dopasowaniu Kołnierza, owiń pas wokół głowy pacjenta tak, aby przechodził przez środek czoła.
3. Dociągnij oba pasy jednocześnie z jednakową siłą nacisku i naprężenia. Przewiń paski przez klamry i zapnij.
4. Sprawdź, czy wszystkie elementy stabilizujące przylegają do pacjenta i czy pacjent przyjął pozycję neutralną, prawidłową.



Dodatkowe rzepy dołączone do zestawu należy przykleić do kołnierza i na opaskę HALO w celu lepszego mocowania.

ZDEJMOWANIE:

1. Przed zdjęciem HALO i Kołnierza, zaznacz położenie pasa z rzepem, by zapewnić właściwe dopasowanie przy ponownym zakładaniu.
2. Przytrzymaj HALO i Kołnierz, rozluźnij pas HALO z prawej a następnie z lewej strony.
3. Aby usunąć Kołnierz, rozepnij rzep z prawej, a następnie z lewej strony.
4. Zdejmij z pacjenta HALO i kołnierz.

ROZMIAR: UNIWERSALNY



OSTATECZNĄ DECYZJĘ DOTYCZĄCĄ ZASTOSOWANIA I WYBORU ZAOPATRZENIA PODEJMUJE LEKARZ SPECJALISTA.

INSTRUCTIONS FOR USE

COLLAR STARK HALO – Halo brace

The COLLAR STARK HALO stabilizer for adults is intended to be used with COLLAR STARK only. Using the stabilizer properly is important for the patient immobilization and comfort. The two-part structure consists of the front part and the back part that are included in the set.

PUTTING ON:

1. Adapt the Halo Stabilizer to the plastic reinforcement in the back part of the collar by inserting it from the top and attaching the stabilizer to the occiput supporting element. Align both parts.
2. Having adjusted the Collar properly, wrap the strap around the patient's head over the centre of the forehead.
3. Adjust both straps at the same time using the same pressure and tension force. Pass the straps through the buckles and fasten them.
4. Make sure that all the stabilizing elements fit closely to the patient and that the patient has adopted a natural and correct posture.



Additional hook-and-loop fasteners included in the set are to be attached to the collar and the HALO band for better fixing.

TAKING OFF:

1. Before taking the HALO and Collar off, mark the hook-and-loop strap position to ensure proper fit when you put them again.
2. Hold the HALO and Collar, loosen the HALO strap at the right and then at the left side.
3. In order to take the Collar off, unfasten the hook-and-loop fastener at the right and then at the left side.
4. Remove the HALO and Collar from the patient.

SIZE: UNIVERSAL



THE FINAL DECISION REGARDING THE APPLICATION AND SELECTION OF EQUIPMENT IS MADE BY A SPECIALIST PHYSICIAN.